



X-Rite Europe GmbH  
Althardstrasse 70  
CH-8105 Regensdorf  
Switzerland

Phone: +41 44 842 24 00  
Fax: +41 44 842 22 22

# Certificate of Performance

**Model:** Eye-One

**Serial No:** 3.278-340708

This notification serves to certify that the unit described above has been inspected and tested to be in accordance with the specifications published by X-Rite.

## Specifications:

Inter instrument agreement:	Average 0.4 $\Delta E^*_{94}$ , max. 1.0 $\Delta E^*_{94}$ (deviation from X-Rite manufacturing standard at a temperature of 23°C for single measurement mode on 12 BCRA tiles (D50, 2°))
Short term repeatability:	0.1 $\Delta E^*_{94}$ (D50, 2°), mean value of 10 measurements every 3 seconds on white
White reference:	Ceramic tile traceable to NIST (National Institute for Standards and Technology, USA). Traceability is established using Certified Reference Material with serial number NIST CS-10

**Next Calibration Due Date:** 4/19/2011

Regensdorf, 19. January 2010

Rainer Hövel  
Manager NPI, ME and Quality

**Headquarters:**

4300 44th Street, Grand Rapids, MI 49512, USA,  
Phone: 800-248-9748 or +1 616-803-2000

**European Headquarters:**

Althardstrasse 70, 8105 Regensdorf, Switzerland,  
Phone: +41 44 842 24 00, Fax: +41 44 842 22 22

**Asia-Pacific Headquarters:**

Room 808-810 Kornhill Plaza – Office Tower, No. 1 Kornhill Road, Quarry Bay, Hong Kong, China,  
Phone (852) 2568-6283, Fax (852) 2885-8610